Calibration services Ensure compliance with minimal process interruptions





Calibration services

Endress+Hauser is the only process instrument manufacturer with the ability to provide accredited calibrations in both the laboratory and on-site for our own and other manufacturers' products.

Calibration is a necessary action that ensures your instrument measurements are accurate. For some applications in the process industries, periodic calibration to the National Institute of Standards and Technology, or NIST, standards are a requirement to comply with quality and safety standards. We can help you stay in compliance while reducing your costs and increasing the process uptime.

Accreditation is the formal recognition of an organization's technical competence in performing specific services such as calibration. This recognition is issued by authorized

bodies, often a national metrology authority working in strict compliance with comprehensive international codes of practice.

Endress+Hauser performs instrument calibrations across a variety of measuring principles. We provide accredited calibrations in both our Greenwood, Indiana and Houston, Texas laboratories and accredited on-site calibration with our mobile calibration rigs. We even extend our calibration services to third-party equipment and some non-traditional flow disciplines to reduce time, effort and cost in terms of coordination and documentation.

```
Ø
```

Calibration services for Flow, pressure and tempertaure



Calibration Measurement Parameters

Parameter	Equipment type	Calibration location
Temperature	 Resistance thermometer Probe, temperature transmitter Probe, display 	On-site In the laboratory
Pressure	ManometerPressure sensorPressure transmitter	On-site In the laboratory
Flow	 Electromagnetic flowmeter Vortex flowmeter Coriolis flowmeter Ultrasonic flowmeter Thermal flowmeter Mechanical flowmeter 	On-site In the laboratory
Conductivity	Conductivity measuring chain including cell, transmitter and cable	On-site In the laboratory
рН	pH measuring chain including cell, transmitter and cable	On-site In the laboratory
Other parameters	On request, we can calibrate other par your local Endress+Hauser partner for	rameters. Please contact further information

Accredited facilities

Thanks to accredited laboratories installed in our factories and our service organization, we can offer the accredited calibration of:

- Flow
- Pressure
- Temperature measurement devices
- Voltage, current

On-site calibration

Having also invested in mobile reference tools and in the deployment of competent service organizations in many countries, Endress+Hauser can also perform on-site calibration — ask your local Sales Center about any specific requests.



Quality calibration for your process lifecycle

Thanks to our numerous calibration laboratories and primary facilities, we have carried out more than one million calibrations.

Above all, we see our calibration service as part of your repair and maintenance planning. Our aim is to provide you with complete calibration management solutions. At the forefront of this is the tuning of calibration cycles, planning dates for carrying out calibration work, the coordination of personnel and certified calibration equipment.

1 **Calibration consulting**

Endress+Hauser will help you to assess your metrology plan by establishing your calibration specifications per parameter (maximum permissible errors, periodicities) or by defining the right reference tools according to their uncertainty. Together, we will work out which parameters have to be calibrated on-site and, in the case of high accuracy requirements, which instruments need laboratory calibration.

Calibration SOPs

We offer a full range of Standard Operating Procedures to support our on-site work. SOPs ensure that our work is repeatable all over the world. We also provide site specific SOP's noting which measuring parameters have to be calibrated on-site and in case of high acccuracy requirements, which need calibrating in the laboratory.

Test equipment 3

Local service centers provide a one-stop calibration and repair service to a wide range of test, measurement and process control instrumentation. So whether you need pressure, analytical, temperature or flow calibration, look no further. All of our facilities are traceable to national and international standards. This means you are guaranteed the highest level of service compliant with the ISO 17025 standard.

Trained employees

4 On-site calibration is performed by specialist, highly trained staff. This relieves your in-house maintenance staff from routine time-consuming tasks and allows them to focus on improving plant availability. For you this means reliable advice, optimum performance of your instruments and true cost-effectiveness.



Calibration work 5

Our primary calibration facilities operate and are accredited to ISO 17025 and are located around the world. We own and operate more calibration laboratories than any other instrumentation supplier. As a leading supplier of field instrumentation we not only can calibrate, but also quickly and efficiently adjust, repair or replace equipment that is failing to meet the specified criteria. Our specialists have the necessary skills and equipment to calibrate all makes of instruments.

Calibration documentation 6

We support our service with certified and traceable documentation. A calibration certificate compliant with the ISO 17025 standard is issued. It details all required data in form that is easy to understand. Importantly, it also satisfies all relevant authorities.

Calibration and Life Cycle Management software Custom software tool that helps to control the scheduling activity around your installed base, providing traceable and auditable records.

Laboratory and on-site accredited calibrations

Endress+Hauser provides accredited calibrations in our laboratories in Greenwood, Indiana and Pearland, Texas. Accredited on-site calibration is also performed nationwide, using our mobile calibration rigs.

Calibration ensures that instrument measurements are accurate. For some processes, periodic calibration to NIST standards is a requirement to comply with quality and safety standards. We can help you stay in compliance with reducing your costs and increasing process up-time.

As a customer of Endress+Hauser, you can be sure that we are thoroughly familiar with all your needs regarding every aspect of calibration and measuring accuracy.

ISO 17025 ISO/IEC 17025 is an internationally accepted standard, covering "general requirements for the competence of testing and calibration laboratories." ISO/IEC 17025 outlines the stringent requirements calibration facilities must meet to achieve and keep the accreditation which includes demonstrating that they operate a quality system, are technologically competent and are able to generate accurate results.

The Calibration Certificate

Often a meter will need an official calibration certificate, i.e. flow metering in regulated industries or for volume measure in large-bore water pipelines. For this reason, Endress+Hauser is an officially accredited calibration provider for the measured variable "flow" and performs Experience in many industries over many years is reflected in our calibration service. Calibration is performed on Endress+Hauser instruments and various other manufacturers' as well.

What is accreditation

Accreditation is the formal recognition of a body's technical competence in performing specific services such as calibration. This recognition is issued by authorized bodies, often a national metrology authority working in strict compliance with comprehensive international codes of practice.

Scope of Accreditation The scope of accreditation includes the principles of flow, pressure and temperature for both laboratory and on-site calibration. Also included within the scope are electrical parameters associated with current, voltage, resistance and frequency pertaining to internal laboratory calibration. Please refer to the Scope of the Accreditation for specific measuring capabilities by visiting <u>a2la.org</u> (search "accredited organizations" for "Endress+Hauser").

calibration with A2LA, SCS or CNAS certification. These certificates are accepted in all ISO member states and consequently are invariably recognized and accepted by national authorities and in quality audits

Flow, Pressure and Temperature	Accredited Certificate	Standard Certificate
Accrediting Body ILAC Seal	v	
Official Stamp of the National Authority/Calibration Service	v	
Details of the Meter/Customer	v	v
Initial Calibration of Recalibration	v	
All Results of Measurement	v	v
Additional Information About the Calibration References	v	
Traceability and Measuring Uncertainty	v	
Stamp/Signature Calibration Provider/Operator	v	v
Full Scale Value	v	v
Details on Calibration Measurement	v	v

		Official stamp of that Endress+H accordance with compliance with	auser is a lauser is a h ISO/IEC h the app	ional authority/calit iccredited as a calibi 17025 and underta licable standards.	ation provider in akes calibration in	_		
Accrediting body is	Certificate Flow Calibration with Adjustment			Endress+ People for Pro				
a member of ILAC (International Laboratory Accreditation Cooperation)		America Calibration J	an Associat	ion of Laboratory Accred 4 according to ISO/IEC 17025	itations, Cert. No 3041.02			
	4017537870 Purchase order number	ITED		FCP-7.1.6 US / gravi Calibration rig/Method	metric	Full scale or custor	e value (standard mer-specific)	
tails of the	Endress + Hauser Serv/Rep Customer US-3601529502-300 / End	ress+Hauser Inc		520 lb/min Calibrated full scale	(≜ 100%) 7			
ter/customer	Order Nº/Manufacturer PROMASS 83 F DN50 / 2"	less muser me.		Calibrated output 1.8933		Details o measure	on calibration ement	
	AC02FD16000 / AC02FD1 Serial N°	6000		-10 Zero point		Measuri	ng uncertainty of	
ial calibration ecalibration	ASSET NO. 10180069 Tag N° Recalibration			70.5 °F ±0.07 °F Water temperature 0.049 %		calibratio of the m	on rig plus reprodu neter tested	cibility
results of measurement	Condition R0 How Bit 20 103.752 30.2 52.14 20.0 103.752 30.2 52.42 35.3 183.598 30.2 20.26 50.0 260.099 30.2 103.71 74.0 389.701 30.1 155.75 100.2 521.122 30.2 261.04	m meas. Δ ο.r.* IN IN 6 52.160 0.028 3 52.200 -0.007 8 92.242 -0.028 2 130.747 0.027 11 195.609 -0.007 14 261.936 -0.009	Outp.** [ma] 7.19 7.20 9.65 12.01 15.99 20.03 20.04	Expanded uncertainty of measu Measured error % o.r. Totermore limit: s0.1% o.	ement			
			-	-1 -				
ditional information but the rig, traceability d measuring uncertainty	**-c-of making ***Calculated value (4 - 20 m.) For detailed data concerning output specific Calibration, traceable to NIST, has been co The reported expanded uncertainty of mes a normal distribution corresponds to a cow of the reference, the calibration method, th	cations of the unit under test, mpleted on the instruments usurement is stated as the stan trage probability of approx. Os ee environment and of the dev	- - - see Technical Ini used in the equipr dard uncertainty %. The expander vice under test. A	-1- -1-5-0 10 20 30 40 ormation (TI), chapter Performance teric tailbration process (see attacher of measurement multipiled by the cc uncertainty of measurement includity typical value is used for the uncertainty	50 60 70 80 90 100 How How characteristics. Il bitring). weage factor k=2, which for se uncertainty components nty component of the device	a		
ditional information out the rig, traceability d measuring uncertainty mp / Signature ibration provider / erator	⁴ C.c. of reading ⁴ C.c. of reading ⁴ Colculated value (4 - 20 mA) For detailed data concerning output specifi Calibration, raceable to NIST, has been co The reported expanded uncertainty of mea- normal distribution corresponds to a cow of the reference, the calibration method, th	cations of the unit under test, mpleted on the instruments us surement is stated as the stam, range probability of approx. 95 e environment and of the dev	- - - see Technical Ini used in the equipr dard uncertainty %. The expanded rice under test. A	-1- -1- -1- -1- -1- -1- -1- -1- -1- -1-	50 00 70 80 90 100 Pow Bow characteristics. Illining). verage factor k=2, which for suncertainty components nty component of the device	Standa	ard Certific	ate
litional information ut the rig, traceability measuring uncertainty mp / Signature bration provider / erator	*c.c of reading **Calculated value (4 - 20 mA) For detailed data concerning output specific Calibration, raceable to NIST, has been co The reported expanded uncertainty of mea- anormal distribution corresponds to a cow of the reference, the calibration method, if	cations of the unit under test, mpleted on the instruments us surement is stated as the stann range probability of approx. 95 e environment and of the dev	see Technical Ini seed in the equipr dard uncertainty %. The expanded vice under test. A	The second secon	50 00 70 80 90 100 Pew characteristics. Iliting). verage factor k=2, which for suncertainty components nty component of the device	Standa	ard Certific Endress+Hai Posteller Process A	cate
ditional information ut the rig, traceability I measuring uncertainty mp / Signature bration provider / erator	Craction of the second se	cations of the unit under test, mpleted on the instruments u surement is stated as the stam, rarge probability of approx. 95 exervitornment and of the dev	- - - see Technical Im seed in the equipt dard uncertainty %. The expanded rice under test. A	Operator Coperator C	50 60 70 80 90 100 Bew characteristics. Iliting). werage factor k=2, which for succertain recomments nty component of the device	Standa	ard Certific Endress+Hai Regist for Process A	ate user दि
litional information ut the rig, traceability measuring uncertainty mp / Signature bration provider / rrator	- - - **Cacionality - - **Contained value (4 - 20 mA) - - For detailed data concerning output specific Concerning output specific - Calibration, raceable to NIST, has been control and the reference, the calibration method, the control of the reference, the calibration method, the calibration - Date of calibration - - - This certificate shall not be published or reaccept with the prior written approval of the calibration -	cations of the unit under test, mpleted on the instruments us surement is stated as the stam, rarge probability of approx. 95 e environment and of the dev	see Technical Im seed in the equipt dard uncertainty %. The expander rice under test. A	Operator Coperator Coperator Coperator Coperator Endress Hauser Inc. 2350 Endress Hauser Inc. 2350 Endress Place Greenwood, IN 46143	50 60 70 80 90 100 Bew characteristics. 11 listing). werage factor k=2, which for suncertainty components nty component of the device Flow Calibration with wereast names 3003764016 Pusher over number US-300407533.100 / Endress-1 Other V Numberson	Adjustment	Ard Certific Endress+Hai Regel for Process A People for Process A St. 1557 Ib/min Calibration International	Cate User Œ (≙100%)
itional information ut the rig, traceability measuring uncertainty mp / Signature bration provider / rator	- - - - - - ***Cacinated value (4 - 20 m.) For detailed data concerning output specific Calibration, recarable to NIST, has been co The reported expanded uncertainty or measuremost on a cow of the reference, the calibration method, the Date of calibration This certificate shall not be published or reexcept with the prior written approval of the	- - - <td>see Technical Im used in the equip dard uncertainty %. The expandence vice under test. A</td> <td>Operator Operator Operator Corrention (I) Adapter Performance beta calibration process (see attaches of measurement multiplied by the co- uncertainty of measurement includ uppical value is used for the uncertainty Operator Endress-Hauser Inc. Endress-Hauser Inc. Greenwood, IN 46143</td> <td>50 60 70 80 90 100 Prov characteristics. Thisting). werage hears k=2, which for suncertainy components ny component of the device Flow Calibration with www.smass 3003764016 Transae dom number US 3004907533-100 / Endress- dom VMandatare 83F50-P1U1/101 Other With Australiant Barbard and Far</td> <td>Adjustment</td> <td>Archevel Archev</td> <td>cate</td>	see Technical Im used in the equip dard uncertainty %. The expandence vice under test. A	Operator Operator Operator Corrention (I) Adapter Performance beta calibration process (see attaches of measurement multiplied by the co- uncertainty of measurement includ uppical value is used for the uncertainty Operator Endress-Hauser Inc. Endress-Hauser Inc. Greenwood, IN 46143	50 60 70 80 90 100 Prov characteristics. Thisting). werage hears k=2, which for suncertainy components ny component of the device Flow Calibration with www.smass 3003764016 Transae dom number US 3004907533-100 / Endress- dom VMandatare 83F50-P1U1/101 Other With Australiant Barbard and Far	Adjustment	Archevel Archev	cate
ditional information but the rig, traceability I measuring uncertainty mp / Signature bration provider / erator	- - - - - - **Cacintration statiled data concentrating outury specific Calibration, reached to NIST, has been co The reported expanded uncertainty of mean anomal distribution corresponds to a cow of the reference, the calibration method, the calibration method, the calibration Date of calibration This certificate shall not be published or reaccept with the prior written approval of the calibration	reduced other than in full, he issuing laboratory.	e Technical Im seed in the equip dard uncertainty %. The expandence rice under test. A	Operator Operator Endress-Hauser Inc. Softward, IN 46143	50 60 70 80 00 100 Prov characteristics. Iliting). werage heator k=2, which for successful records and the device Flow Calibration with were the theory of the device Flow Calibration with were the theory of the device Successful records and the device successful records and the device Successful records and the device successful records and the device Successful records and the device successful records and the	Adjustment	ECP-7.1.0 US Endress + Hai Negele for Process A ECP-7.1.0 US Calibration ng 55.1.557 (b):/mln Calibration Interface Calibration In	cate User III (≙100%)
itional information ut the rig, traceability measuring uncertainty mp / Signature bration provider / rator	- - - - - - <td>reduced other than in full, te issuing laboratory.</td> <td>see Technical Ini ased in the equip dard uncertainty %. The expandence rice under test. A Page 1 / 2</td> <td>Operator Endress-Hauser Inc. Corrention (II) A chapter Ferformance here calibration process (see attaches of messurement multiplied by the co- tuncertainty of messurement licitid typical value is used for the uncertainty Operator Endress-Hauser Inc. Greenwood, IN 46143</td> <td>50 00 70 00 90 100 Pew characteristics. Ilining). wreage factor k=2, which for suncertainty components ny component of the device Flow Calibration with www.sharast 3003764016 Pactare offer maler US-3004097533-100 / Endress- torse // Number US-3004097533-100 / Endress- torse // Number 3850-P1U/101 Offer cols 3850-P1U/101 Offer cols 3850-P</td> <td>Adjustment Hauser Inc.</td> <td>ACCOMPANIES AND ADDRESS AND AD</td> <td>(≜ 100%)</td>	reduced other than in full, te issuing laboratory.	see Technical Ini ased in the equip dard uncertainty %. The expandence rice under test. A Page 1 / 2	Operator Endress-Hauser Inc. Corrention (II) A chapter Ferformance here calibration process (see attaches of messurement multiplied by the co- tuncertainty of messurement licitid typical value is used for the uncertainty Operator Endress-Hauser Inc. Greenwood, IN 46143	50 00 70 00 90 100 Pew characteristics. Ilining). wreage factor k=2, which for suncertainty components ny component of the device Flow Calibration with www.sharast 3003764016 Pactare offer maler US-3004097533-100 / Endress- torse // Number US-3004097533-100 / Endress- torse // Number 3850-P1U/101 Offer cols 3850-P1U/101 Offer cols 3850-P	Adjustment Hauser Inc.	ACCOMPANIES AND ADDRESS AND AD	(≜ 100%)
ditional information but the rig, traceability d measuring uncertainty mp / Signature ibration provider / erator	- - - - - - - - - - - - - - - - - - - - - - - - - - - - - - - **Cation (interpreted expanded uncertainty of new anomal distribution corresponds to anomal distribution corresponds or anomal distribution corresponds or anomal distribution corresponds or anomal distribution method, the reference, the calibration This certificate shall not be published or a reactive with the prior written approval of the except with the prior written approval of the distribution of the distribution of the distribution of the distribution distributico distributico distribution distributico distrib	reduced other than in full, be issuing laboratory.	see Technical Ini ased in the equip dard uncertainty %. The expandence rice under uss. A Page 1 / 2	Operator Control of the second secon	50 00 70 80 90 100 I) Rev characteristics. Illining). Illining). Illining). versage factor k=2, which for summariant or summariant of the device Illining). Illining). versage factor k=2, which for summariant or summaria			:ate User (± 100%)

Laboratory calibrations

Our laboratory calibration facilities are located in Greenwood, Indiana (near Indianapolis), and Pearland, Texas (near Houston). These facilities house state-of-theart calibration tools in order to provide the most accurate



calibration results. We work in collaboration with our production centers and our authorized service providers to ensure reliable outcomes for our customers.



Greenwood, Indiana

Calibration Capabilities



Flow (Liquid)

Principle	Min. MPE (o.r.)	Size	Uncertainty (o.r.)	Certificate
Coriolis	±0.1%	1⁄2" to 10" (DN15 to DN250)	±0.05%	1
	±0.1%	1/24" (DN01)	±0.05%	
Coriolis – Premium*	±0.015%/0.05%	1/12" to 10" (DN02 to DN250)	±0.015%*	2
Electromagnetic	±0.2%	½" to 12" (DN15 to DN300)	±0.1%	1
	±0.2%	1/12" to 42" (DN02 to DN 1050)	±0.1%	2
Vortex	±0.75%	½" to 12" (DN15 to DN300)	±0.25%	1
Ultrasonic – In-line	±0.5%	1" to 6" (DN25 to DN150)*	±0.2%	1
	±0.5%	14" to 48" (DN350 to DN1200)	±0.2%	2
Ultrasonic – Clamp On	±2%	2" to 4" (DN0 & DN100)	±0.5%	Verification only

*Coriolis – Premium calibration can only be executed on Endress+Hauser flowmeters

MPE = Maximum Permissible Error

o.r. = of reading

1. Certificate #3041.02 - Service Center Greenwood - All manufacturers

2. Certificate #1897.01 - Production Center Greenwood - Endress+Hauser flowmeters only

Flow Range

Location	Туре	Min [*]	Max*
Service Center	Mass Flow Rate	13.2 lb./m (0.1 kg/s)	25,133 lb./m (190 kg/s)
	Volume Flow Rate	1.58 gal./m (0.1 l/s)	3,021 gal./m (190 l/s)
	Low Flow Rate	8 ml/min	100 ml/min
Production Center	Mass Flow Rate	2.65 lb./m (0/02 kg/s)	87,303 lb./m (660 kg/s)
	Volume Flow Rate	0.32 gal./m (0.02 l/s)	10,493 gal./m (660 l/s)

*Minimum and maximum flow ranges are line size dependent

Flow (Air)

Uncertainty	≤± 0.3%
Flow Rates	0-8265 lb./hour (0-3750 kg/hour)
Pressure	± 0.72 psia (±50 mbarA)
Temperature	≤± 0.54 °F (≤± 0.3 °C)
Relative Humidity	≤45%
Line Sizes	$\frac{1}{2}$ " to 4" (inline devices) DN15- DN100; ≥4" (insertion devices) ≥DN100
Instruments Support	T-mass A150, t-mass B 150 t-mass 65F, t-mass 65I

Pressure

Principle	Min. MPE	Size	Uncertainty
Gauge – Pneumatic	±0.040% o.f.s.	-13.1 to 1500 psig (130 bar)	±0.013% o.f.s.
Absolute	±0.040% o.f.s.	1.6 to 1515 psia (110 mbarA to 104 barA)	±0.013% o.f.s.
Differential	±0.040% o.f.s.	-15 to 1500 psid (-1 to 103 bard)	±0.013% o.f.s.
Gauge – Hydraulic	±0.040% o.f.s.	600 to 10,000 psi (41 to 689 bar)	±0.013% o.f.s.

Temperature

Principle	Min. MPE	Size	Uncertainty
Transmitter with RTD/ Thermocouple"	±0.45 °F (0.25 °C)	-22 to 392 °F (-30 to +200 °C)	±0.06 °F (0.03 °C)

o.f.s= of full scale

MPE= Maximum Permissible Error

Pearland, Texas

Calibration Capabilities



Flow (Liquid)

Principle	Min. MPE (o.r.)	Size	Uncertainty (o.r.)	Certificate
Coriolis	±0.1% o.r.	1/12" to 10" (DN02 to DN250)	±0.05% o.r.	3
Coriolis – Premium*	±0.05% o.r.	1 1/2" to 3" (DN40 to DN80)	±0.015% o.r.	3
Electromagnetic	±0.2% o.r.	1/8" to 12" (DN04 to DN300)	±0.1% o.r.	3
Vortex	±0.75% o.r.	½" to 4" (DN15 to DN100)	±0.25% o.r.	3
Ultrasonic – In-line	±0.5% o.r.	1" to 6" (DN25 to DN150)	±0.2% o.r.	3
Ultrasonic – Clamp On	±2% o.r.	2", 4", & 6" (DN50, DN100, & DN150)	±0.5% o.r.	Verification only

*Coriolis – Premium calibration can only be executed on Endress+Hauser flowmeters o.r. = of reading

3. Certificate #3041.01- Service Center Pearland – All Manufacturers

Flow Range

Location	Туре	Min*	Max*
Service Center	Mass Flow Rate	0.1097 lb./m (0.00083 kg/s)	17,196 lb./m (130 kg/s)
	Volume Flow Rate	0.01316 gal./m (0.00083 l/s)	2060 gal./m (130 l/s)

*Minimum and maximum flow ranges are line size dependent

Pressure

Principle	Min. MPE	Size	Uncertainty
Gauge – Pneumatic	±0.040% o.f.s.	-14.5 to 600 psig (41.3 barg)	±0.013% o.f.s.
Absolute	±0.040% o.f.s.	0.13 to 615 psia (8.9 mbarA to 42 bara)	±0.013% o.f.s.

Temperature

Principle	Min. MPE	Size	Uncertainty
Transmitter with RTD/	±0.45 °F (0.25 °C)	-22 to 392 °F (-30 to +200 °C)	±0.06 °F (0.03 °C)
Thermocouple"			

o.f.s= of full scale

MPE= Maximum Permissible Error



Nationwide on-site calibration

Endress+Hauser, in collaboration with its Authorized Service Providers, executes on-site calibration by using the state-of-the-art calibration technologies installed in our mobile calibration rigs. On-site calibrations can take place across the nation, through our elaborate network of certified calibration technicians and over 20 mobile calibration rigs.

On-Site Calibration Capacities

Flow (Liquid)

Principle	Min. MPE	Equipment/ Range	Uncertainty
Mass (Coriolis)	±0.25% o.r.	Portable In-line care, Line Size 1/12"-3" (DN02-DN80) Standard 2-point calibration	±0.12% o.r.
Volumetric (Magmeter & Vortex)	±0.5% / ±1.5% o.r.	Portable Mobile Rig, Line Size 3/8"-2" (DN08- DN50) Standard 2-point calibration	±0.17% o.r.
Flow Verification	With Heartbeat T	echnology [®] or Ultrasonic Clamp-on (liquid only)	

Pressure and Temperature

Principle	Min. MPE	Equipment/ Range	Uncertainty
Pressure	±0.14% o.f.s.	With Mobile Calibrations –15 to 5,000 psig (-1 to 345 bar) Standard 1- or 2-point calibration	±0.044% o.f.s.
Temperature	±0.48°F	With Mobile Calibrators 23 to 707 °F (-5 to 375 °C) Standard 1- or 2-point calibration	±0.16 °F

MPE= Maximum Permissible Error

o.r.= of reading

o.f.s.= of full scale

- 1. Accuracy statement based on best uncertainty of reference equipment test (UUT) with a calibration measurement capacity of at least 2:1
- 2. Depending on customer's capabilities
- 3. Pressure minimum MPE is dependent on range, may be lagrer in draft ranges
- 4. Temperature minimum MPE is based on oil bath, may be larger in dry block

Fulfilling the needs of all industries

Does the instrumentation that controls your critical quality processes need regular checking, validation and calibration? Do you need cost-effective service that is fast, high quality, traceable and accredited? Do you need clear and concise calibration certificates?

Endress+Hauser covers all these critical aspects and can perform and advise on all aspects of calibration from in-situ testing to sully accredited factory calibration. We calibrate your equipment at just the right time for you. This ensures optimal process performance at minimum cost.



Oil & gas industries

- On-site flow calibration rigs up to 2" (3" in-line)
- Laboratory calibration of all kinds of custody transfer devices
- Laboratory calibration of flowmeters up to 48"

Enviromental industry

- Verification of flowmeters
- Laboratory calibration of custody transfer instruments
- Laboratory calibration of flowmeters up to 48"

Life sciences industry

As a compliant partner, we:

- Perform calibration of process measuring points according to cGMP regulations and universal guidelines
- Document any service provided
- Prove traceability by ensuring the conformity to Standard Operating Procedures
- Keep service staff trained to GMPs and SOPs.

Food & beverage industries

- Calibration within regulatory requirements
- Calibration management strategies to support production costs concepts

Chemical industries

- On-site flow calibration rigs up to 2" (3" in-line)
- Perform calibration of process measuring points
- Laboratory calibration of flowmeters up to 48"

Renewable fuels industry

On-site flow calibration rigs up to 2" (3" in-line)

How to get started with Endress+Hauser calibration services

It's easy!



Choose any of the following options to get started:



Contact your local sales representative





Visit Calibration services for Flow, pressure and tempertaure webpage



- Fill out our online Calibration Inquiry Form
- Download a calibration data sheet

CP01428H/60/EN/02.22



www.addresses.endress.com